



CheKine™ Micro Tannin Assay Kit

Cat #: KTB1541

Size: 48 T/48 S

96 T/96 S

	Micro Tannin Assay Kit		
REF	Cat #: KTB1541	LOT	Lot #: Refer to product label
	Detection range: 0.0156-1 mg/mL		Sensitivity: 0.0078 mg/mL
	Applicable samples: Plant Tissues, Liquid samples such as Juice and Honey		
	Storage: Stored at 4°C for 6 months, protected from light		

Assay Principle

Tannin is a kind of polyphenolic compounds widely present in plants, also known as plant polyphenols. It generally has astringent taste, can precipitate proteins, alkaloids, and polysaccharides, and can have complex or electrostatically interact with various metal ions. According to chemical structure, tannin can be divided into hydrolyzable tannin and condensed tannin. The ability of tannin binding to proteins is also known as astringency. Its astringency is the basis of various physiological activities, such as hemostasis, inhibition of microorganisms, anti-allergy, anti-mutation, anti-tumor, anti-aging and other physiological activities, and it is also one of the factors affecting the taste of products. CheKine™ Micro Tannin Colorimetric Assay Kit provides a convenient tool for detecting the tannin content such as plant tissue, liquid samples such as juice and honey. The principle is that tannin reacts with phosphomolybdic acid in an alkaline environment to form a blue compound, which has a characteristic absorption peak at 760 nm. The tannin content of the sample can be calculated by measuring the absorbance at 760 nm.

Materials Supplied and Storage Conditions

Kit components	Size		Storage conditions
	48 T	96 T	
Assay Buffer	5 mL	10 mL	4°C
Chromogen	5 mL	10 mL	4°C, protected from light
Standard	Powder×1 vial	Powder×1 vial	4°C, protected from light

Note: Before formal testing, it is recommended to select 2-3 samples with large expected differences for pre-experiment.

Materials Required but Not Supplied

- Microplate reader or visible spectrophotometer capable of measuring absorbance at 760 nm
- 96-well plate or microglass cuvette, precision pipettes, disposable pipette tips
- Incubator, water bath, centrifuge
- Deionized water
- Homogenizer (for plant tissue with less fibers)
- Oven, pulverizer or wall breaker, 40-mesh sieve (for plant tissue with more fibers)

Reagent Preparation

Assay Buffer: Ready to use as supplied. Equilibrate to room temperature before use. Store at 4°C.

Chromogen: Ready to use as supplied. Equilibrate to room temperature before use. Store at 4°C, protected from light.

Note: Chromogen has certain irritation, so personal protection is recommended during use.

Standard: Add 1 mL deionized water to dissolve before use. The concentration is 5 mg/mL. Store at 4°C, protected from light.

Standard curve setting: Dilute 5 mg/mL Standard with deionized water to 1, 0.5, 0.25, 0.125, 0.0625, 0.0313, 0.0156 mg/mL standard solution as shown in the table below.

Num.	Volume of Standard	Volume of Deionized Water (μL)	The Concentration of Standard (mg/mL)
Std.1	200 μL of 5 mg/mL	800	1
Std.2	100 μL of Std.1(1 mg/mL)	100	0.5
Std.3	100 μL of Std.2 (0.5 mg/mL)	100	0.25
Std.4	100 μL of Std.3 (0.25 mg/mL)	100	0.125
Std.5	100 μL of Std.4 (0.125 mg/mL)	100	0.0625
Std.6	100 μL of Std.5 (0.0625 mg/mL)	100	0.0313
Std.7	100 μL of Std.6 (0.0313 mg/mL)	100	0.0156

Note: Always prepare fresh standards per use; Diluted Standard Solution is unstable and must be used within 4 h.

Sample Preparation

1. Plant tissues

(1) Plant tissue with less fibers: Weigh 0.1 g tissue, add 1 mL deionized water and fully homogenize, then transfer to an EP tube, extract in a water bath at 80°C for 30 min. Centrifuge at 8,000 g for 10 min at 25°C. Use supernatant for assay.

(2) Plant tissue with more fibers: Sample was dried to constant weight, pulverized, passed through a 40-mesh sieve. Then weigh 0.1 g, add 1 mL deionized water mix well, extract in a water bath at 80°C for 30 min. Centrifuge at 8,000 g for 10 min at 25°C. Use supernatant for assay.

2. Liquid samples such as juice and honey: Tested directly.

Assay Procedure

1. Preheat the microplate reader or visible spectrophotometer for more than 30 min, and adjust the wavelength to 760 nm, visible spectrophotometer was returned to zero with deionized water.

2. Add the following reagents respectively into each well of 96-well plate or microglass cuvette:

Reagent	Blank Well (μL)	Standard Well (μL)	Test Well (μL)
Deionized Water	100	90	90
Stds.	0	10	0
Sample	0	0	10
Chromogen	50	50	50
Assay Buffer	50	50	50

3. Mix well, incubate at room temperature (25°C) for 10 min. Then reading the values at 760 nm. Finally, calculate $\Delta A_{\text{Test}} = A_{\text{Test}} - A_{\text{Blank}}$, $\Delta A_{\text{Standard}} = A_{\text{Standard}} - A_{\text{Blank}}$ (only one blank well needs to be detected).

Note: The Blank Well and the Standard Well only need to be done 1-2 times. In order to guarantee the accuracy of experimental results, need to do a pre-experiment with 2-3 samples. If ΔA_{Test} is less than 0.02, increase the sample quantity appropriately. If A_{Test} is greater than 1 mg/mL of $\Delta A_{\text{Standard}}$, the sample can be appropriately diluted with deionized

water, the calculated result multiplied by the dilution factor.

Data Analysis

Note: We provide you with calculation formulae, including the derivation process and final formula. The two are exactly equal. It is suggested that the concise calculation formula in bold is final formula.

1. Drawing of standard curve

With the concentration of the standard solution as the y-axis and the $\Delta A_{\text{Standard}}$ as the x-axis, draw the standard curve. Substitute the ΔA_{Test} into the equation to obtain the y value (mg/mL).

2. Calculate the content of Tannin

(1) By sample weight

$$\text{Tannin (mg/g)} = y \times V_{\text{Extraction}} \div W \times n = \mathbf{y \div W \times n}$$

(2) Calculated by liquid volume

$$\text{Tannin (mg/mL)} = \mathbf{y \times n}$$

Where: $V_{\text{Extraction}}$: deionized water volume added for sample extraction, 1 mL; W: sample weight, g; n: dilution multiple of sample.

Typical Data

Typical standard curve




Figure 1. Standard curve for Tannin.

Examples




Figure 2. Tannin content in rice leaf, corn leaf, grape peel and mango peel respectively. Assays were performed following kit protocol

Recommended Products

Catalog No.	Product Name
KTB1540	CheKine™ Micro Plant Total Phenols (TP) Assay Kit
KTB1530	CheKine™ Micro Plant Flavonoids Assay Kit
KTB1520	CheKine™ Micro Plant Oligomeric Proantho Cyanidins (OPC) Assay Kit

Disclaimer

The reagent is only used in the field of scientific research, not suitable for clinical diagnosis or other purposes. For your safety and health, please wear a lab coat and disposable gloves.